

**1988  
Rationale for Revisions**

**Part I  
Radiation Safety Requirements for Particle Accelerators**

There were no major amendments to Part I for this edition of Volume I (Ionizing Radiation) of the Suggested State Regulations for Control of Radiation (SSRCR). The Technical Review Committee made some editorial changes to Part I for consistency with regulatory language and with the other Parts of the SSRCR (e.g., revision of certain provisions with "and/or" as part of the sentence structure).

Matters for Future Consideration

1. Specifications should be provided for the frequency, accuracy, precision, and traceability of calibrations.
2. In reference to Paragraph I.8(c), whether both the interlocks and their indicators of status should be required to be on the same independent circuits should be reviewed and appropriate changes made.