

## What's NEXT?

The Nationwide Evaluation of X-ray Trends (NEXT) is a national program conducted annually to collect data regarding the x-ray exposure that a standard patient receives for selected x-ray examinations. This program is conducted jointly by the Conference of Radiation Control Program Directors, Inc. (CRCPD), an association of state and local radiation control agencies, and the Food and Drug Administration's (FDA) Center for Devices and Radiological Health (CDRH).

In 2002, the selected examinations were the Abdomen and Lumbosacral (LS) Spine. There are two trifold for the 2002 survey. See the *LS Spine Trifold for LS Spine data*. Two hundred and seventy (270) facilities were surveyed, which also includes data from facilities that use digital-based imaging. A clinically validated patient equivalent phantom was used to capture patient exposure data. This phantom represents an adult of approximately 1.73 m (5 ft 9 in), weighing 75 kg (165 lb.), with an equivalent patient thickness measured P/A, of 23 cm (9 in).

Specific information was obtained pertaining to the X-ray equipment, facility workload, and radiographic technique, as well as information relating to air kerma including, grid use, beam quality, darkroom fog, and the quality of film processing.

Downloadable NEXT survey summaries and further information on NEXT surveys are available from the CRCPD at [www.crcpd.org/NEXT.aspx](http://www.crcpd.org/NEXT.aspx) or from the Food and Drug Administration at [www.fda.gov/cdrh/radhlth/next.html](http://www.fda.gov/cdrh/radhlth/next.html).

*The information contained herein is for guidance. The implementation and use of the information and recommendations are at the discretion of the user. The mention of commercial products, their sources, or their use in connection with material reported is not to be construed as either an actual or implied endorsement by CRCPD or CDRH.*

## SURVEY RESULTS FOR YOUR FACILITY


|                                    | Your Result | Survey Mean |
|------------------------------------|-------------|-------------|
| Entrance Skin Air Kerma (mGy)      |             | 2.7         |
| Tube Potential (kVp)               |             | 75          |
| Processing Speed STEP* Test Result |             | 107         |
| Darkroom Fog (OD)                  |             | 0.06        |
| Phantom Film (OD)                  |             | 1.82        |
| mAs                                |             | 75.5        |
| Low Contrast Objects               |             | 5           |
| High Contrast Objects              |             | 4           |

\*Sensitometric Technique for the Evaluation of Processing

CRCPD, 1030 Burlington Lane, Suite 4B, Frankfort, KY 40601

## Nationwide Evaluation of X-Ray Trends (NEXT)

### 2002 Abdomen Radiography Survey



Conference of Radiation Control Program Directors, Inc.




and

Food and Drug Administration  
Center for Devices and Radiological Health




Pub. August 2006

**Entrance Skin Air Kerma: All Facilities**




**Entrance Skin Air Kerma: Film Screen vs. Digital**




**2002 Abdomen Data Summary**  
(Mean values listed below)

|                                   | HOSP  | PP    | OTR   |
|-----------------------------------|-------|-------|-------|
| Number of Surveys                 | 104   | 46    | 31    |
| ESAK (mGy)                        | 2.7   | 2.7   | 2.9   |
| Tube Current (mA)                 | 424   | 319   | 265   |
| mAs                               | 103   | 64    | 52    |
| Tube Pot. (kVp)                   | 75    | 75    | 77    |
| HVL (mm Al)                       | 3.0   | 3.1   | 3.0   |
| Darkroom Fog (OD)                 | .06   | .07   | .55   |
| Film Proc. Speed                  | 107   | 110   | 107   |
| Phantom Film (OD)                 | 1.76  | 1.92  | 1.84  |
| High/Low Contrast Objects Visible | 4 / 5 | 4 / 5 | 4 / 5 |

**Phantom Film Optical Density: All Facilities**




**Tube Current-Exposure Time Product (mAs): All Facilities**




**Abbreviation Key:**

- Hosp = Hospitals
- PP = Private Practice
- OTR = Other types of facilities

**Film Processing (STEP) \* Speed: All Facilities**



**Darkroom Fog Optical Density: All Facilities**



**Film Screen vs. Digital Imaging Systems: Entrance Skin Air Kerma (ESAK) (mGy) Data: Overall and by Mode**

| System Type       | ESAK | AEC | Manual |
|-------------------|------|-----|--------|
| Film Screen       | 2.6  | 2.4 | 3.4    |
| Digital (CR & DR) | 2.7  | 2.5 | 3.9    |

\* Mammography Quality Standards Act (MQSA) Std. for STEP Speed: 80-120 is considered acceptable